

HemaTechnologies Inc.

HemaTechnologies, Inc.
291 Route 22
Lebanon, NJ 08833 (USA)

Tel: 908 823 9430
Fax: 908 823 9428
E-mail: hematek@hematek.com

ASSAY SHEET

ESR CONTROL

MANUFACTURER R&D SYSTEMS, Inc.
MATERIAL SEDRite™ Plus - Lot # SR-078
EXPIRATION DATE 05 January 2009

		LEVEL I Lot No. SR 078-1		LEVEL II Lot No. SR 078-2	
METHOD	UNITS	MEAN	RANGE	MEAN	RANGE
ESR STAT PLUS	mm/hr	13	7 - 19	70	41 - 99

Note: 1. Follow the enclosed R & D Systems instructions for use.
2. Assay values provided by HemaTechnologies, Inc.

6/19/08